

IHC CDF Dredge Project : Air Monitoring Alarm Response Log

Air Monitoring Station: M1 M2 M3 M4

Cerex Unit (Naphthalene) Thermo Unit (Particles) PID Monitor

Date & Time of alarm: 11/17/2012 7:45 AM

Email sent to site Technician (Yes or No).

Technician responded to the alarm: Scott Peterson

1. Alarm cause by:

- | | |
|-------------------------------------|-----------------------------|
| <input type="checkbox"/> | Loss of Power |
| <input type="checkbox"/> | Loss of Radio Communication |
| <input checked="" type="checkbox"/> | Out of Calibration |
| <input type="checkbox"/> | UV Alignment |
| <input type="checkbox"/> | Blockage in Air Tube |
| <input type="checkbox"/> | Air Quality |
| <input type="checkbox"/> | Other: |

2. Corrective Actions taken (Yes or No).

3. Dredging suspended (Yes or No).

4. Alarm logged in air monitoring action spreadsheet. (Yes or No).

Description of Action Taken:

M4 Thermo Unit alarmed for a high particle readings. The units naphelometer was recalibrated at 9:30 AM 11/17/12 and inspected at 11:00 AM 11/17/12 to confirm proper function.