## $\square$ M1 $\mathbf{X}$ M2 $\square$ M3 $\square$ M4 □ Dock **Air Monitoring Station: Air Monitoring Instrument:** X Thermo Unit (Particles) ☐ PID Monitor ☐ Cerex Unit (Naphthalene) (Yes) or No) Date & Time of alarm: <u>11/21/2012</u> 1:30 AM Email sent to site Technician? Technician responded to the alarm; Scott Peterson (Yes) or No) 1. Was dredging occurring at the time of the alarm? 2. Alarm caused by: Loss of Power Loss of Radio Communication Out of Calibration **UV** Alignment Blockage in Air Tube Air Quality Other: Weather-Heavy fog 3. Corrective Actions taken? (Yes or No

IHC CDF Dredge Project: Air Monitoring Alarm Response Log

(Yes or No)

## Description of Action Taken:

4. Dredging suspended?

5. Alarm logged in air monitoring action spreadsheet?

M2 Thermo unit alarm was caused by heavy fog. The unit was recalibrated at 7:15 AM 11/21/12 and inspected at 12:30 PM 11/21/12 to confirm proper function.