\mathbf{X} M1 \mathbf{X} M2 \square M3 \square M4 □ Dock **Air Monitoring Station: Air Monitoring Instrument:** ☑ Thermo Unit (Particles) ☐ Cerex Unit (Naphthalene) ☐ PID Monitor Date & Time of alarm: 4/06/2013 11:30, 12:15, 1:45, 2:00, 3:15, 3:30 Email sent to site Technician? (Yes) or No Technician responded to the alarm; Scott Peterson (Yes) or No) 1. Was dredging occurring at the time of the alarm? 2. Alarm caused by: Loss of Power Loss of Radio Communication Out of Calibration **UV** Alignment Blockage in Air Tube Air Quality Other: Weather (Yes or(N 3. Corrective Actions taken? 4. Dredging suspended? (Yes or (

IHC CDF Dredge Project: Air Monitoring Alarm Response Log

(Yes or No

Description of Action Taken:

5. Alarm logged in air monitoring action spreadsheet?

Thermo units alarmed from hurricane type winds on Saturday. There were walls of dirt/dust coming from on and off the site. Units were inspected at 4:30 PM 4/6/13 to confirm proper functions. All units were working correctly. Site Project Manager was notified of the alarms.