IHC CDF Dredge Project: Air Monitoring Alarm Response Log $\mathbf{X}\mathbf{M}\mathbf{1}$ \square M2 \square M3 \square M4 □ Dock **Air Monitoring Station: Air Monitoring Instrument:** ▼ Thermo Unit (Particles) ☐ Cerex Unit (Naphthalene) ☐ PID Monitor (Yes) or No) Date & Time of alarm: <u>6/10/2013 6:30 AM</u>, Email sent to site Technician? Technician responded to the alarm; Scott Peterson (Yes) or No) 1. Was dredging occurring at the time of the alarm? 2. Alarm caused by: Loss of Power Loss of Radio Communication X Out of Calibration **UV** Alignment Blockage in Air Tube Air Quality

3. Corrective Actions taken?

(Yes) or No)

4. Dredging suspended?

(Yes or (No))

5. Alarm logged in air monitoring action spreadsheet?

Other: Site Conditions

(Yes) or No)

Description of Action Taken:

A) After check the air monitor it was found that the unit was out of calibration. Inlet tube was taken apart and cleaned. Unit was recalibrated at 9:30 AM 6/10/13 and inspected at 16:00 PM 6/10/13 to confirm proper function.