IHC CDF Dredge Project : Air Monitoring Alarm Response Log						
Air Monitoring Station:	□ M1	□ M2	□ M3	X M4	Dock	
Air Monitoring Instrument:						
Cerex Unit (Naphthalene)			rticles)	□ PID Monitor		
Date & Time of alarm: <u>6/14/2013 10:30 AM</u> ,			Email sent	to site Technician?	(Yes) or No)	
Technician responded to the alarm; Scott Peterson						
1. Was dredging occurring at the time of the alarm?					(Yes) or No)	
2. Alarm caused by:						
Loss of Power						
Loss of Radio Communication						
Out of Calibration						
UV Alignment						
Blockage in Air Tube						
Air Quality						
X Other: Site C	Conditions					
3. Corrective Actions taken?				(Yes) or No)		
4. Dredging suspended?					(Yes or No))	
5. Alarm logged in air monitoring action spreadsheet?					(Yes or No)	

Description of Action Taken:

A) After check the air monitor unit there were no problems found. Unit was recalibrated at 10:00 AM 6/14/13 and inspected at 16:00 PM 6/14/13 to confirm proper function.