## IHC CDF Dredge Project: Air Monitoring Alarm Response Log $\square$ M1 $\square$ M2 $\mathbf{x}$ M3 $\square$ M4 □ Dock **Air Monitoring Station: Air Monitoring Instrument:** X Thermo Unit (Particles) ☐ PID Monitor ☐ Cerex Unit (Naphthalene) (Yes) or No) Date & Time of alarm: <u>8/4/2013 12:00PM</u> Email sent to site Technician? Technician responded to the alarm; Scott Peterson (Yes or No 1. Was dredging occurring at the time of the alarm? 2. Alarm caused by: Loss of Power Loss of Radio Communication Out of Calibration **UV** Alignment Blockage in Air Tube Air Quality Other: Weather Corrective Actions taken? (Yes or (No

## Description of Action Taken:

Dredging suspended?

5. Alarm logged in air monitoring action spreadsheet?

A) No problems were found. Driving to fast causing dust and dry site condictions is believe to have caused the alarms.

(Yes or No