

IHC CDF Dredge Project : Air Monitoring Alarm Response Log

Air Monitoring Station: M1 M2 M3 M4 Dock

Air Monitoring Instrument:

Cerex Unit (Naphthalene) Thermo Unit (Particles) PID Monitor

Date & Time of alarm: 2/5/2014 7:15am-7:45am Email sent to site Technician? (Yes or **No**)

Technician responded to the alarm; Scott Peterson

1. Was dredging occurring at the time of the alarm? (Yes or **No**)

2. Alarm caused by:

- Loss of Power
- Loss of Radio Communication
- Out of Calibration
- UV Alignment
- Blockage in Air Tube
- Air Quality
- Other: weather, cold and wind, snow

3. Corrective Actions taken? (Yes or **No**)

4. Dredging suspended? (Yes or **No**)

5. Alarm logged in air monitoring action spreadsheet? (**Yes** or No)

Description of Action Taken:

A) No problems were found with the air units.