IHC CDF Dredge Project: Air Monitoring Alarm Response Log \square M1 \square M2 \mathbf{X} M3 \square M4 □ Dock **Air Monitoring Station: Air Monitoring Instrument:** ▼ Thermo Unit (Particles) ☐ PID Monitor ☐ Cerex Unit (Naphthalene) Date & Time of alarm: <u>6/05/2014 6:30 PM</u> Email sent to site Technician? (Yes)or No) Technician responded to the alarm; Scott Peterson (Yes or No 1. Was dredging occurring at the time of the alarm? 2. Alarm caused by: Loss of Power Loss of Radio Communication Out of Calibration **UV** Alignment Blockage in Air Tube Air Quality

Description of Action Taken:

4. Dredging suspended?

Other:

3. Corrective Actions taken?

5. Alarm logged in air monitoring action spreadsheet?

o The air intake pump on M3 stopped running due to several power surges. The pump was restarted and no other problems were found.

(Yes or (No

(Yes or (No))

(Yes) or No)