IHC CDF Dredge Project : Air Monitoring Alarm Response Log Air Monitoring Station: □ M1 □ M2 □ M3 □ M4 □ Dock Air Monitoring Instrument:

□ Cerex Unit (Naphthalene) □ Thermo Unit (Particles) □ PID Monitor

Date & Time of alarm: 6/17/2014 Email sent to site Technician? (Yes on No)

Technician responded to the alarm; Scott Peterson

1. Was dredging occurring at the time of the alarm? (Yes or No)

2. Alarm caused by:
□ Loss of Power
□ Loss of Radio Communication
○ Out of Calibration
□ UV Alignment

3. Corrective Actions taken?

Air Quality

Other: Weather

(Yes or No

4. Dredging suspended?

(Yes or (No))

5. Alarm logged in air monitoring action spreadsheet?

Blockage in Air Tube

(Yes) or No)

Description of Action Taken:

- Due to heavy thunderstorms' on 6/16/2014, the Cerex air unit radio lost communications with the server radio. Once the deficiency was noticed, the server was updated with data and the unit was recalibrated on 6/18/2014.
- o No email was sent to the onsite technician to check the unit.