IHC CDF Dredge Project: Air Monitoring Alarm Response Log \square M1 \square M2 \mathbf{X} M3 □ Dock **Air Monitoring Station:** \square M4 **Air Monitoring Instrument:** ▼ Thermo Unit (Particles) ☐ Cerex Unit (Naphthalene) ☐ PID Monitor (Yes) or No) Date & Time of alarm: <u>8/25/2014 2:45AM</u> Email sent to site Technician? Technician responded to the alarm; Scott Peterson (Yes or No 1. Was dredging occurring at the time of the alarm? 2. Alarm caused by: Loss of Power Loss of Radio Communication Out of Calibration **UV** Alignment Blockage in Air Tube Air Quality Other: weather

Description of Action Taken:

4. Dredging suspended?

3. Corrective Actions taken?

5. Alarm logged in air monitoring action spreadsheet?

A. There were thunderstorms around the time of the alarm. Unit was checked on 5/25/2014. No problems were found wrong with the air unit. The units vacuum pump was removed the carbon veins were checked. The units was recalibrated and checked again at 4:00 PM.

(Yes or No

(Yes or No)
(Yes or No)