

IHC CDF Dredge Project : Air Monitoring Alarm Response Log

Air Monitoring Station: M1 M2 M3 M4 Dock

Air Monitoring Instrument:

Cerex Unit (Naphthalene) Thermo Unit (Particles) PID Monitor

Date & Time of alarm: 1/4/2015 4:45-5:30pm Email sent to site Technician? ((Yes) or No)

Technician responded to the alarm; Scott Peterson

1. Was dredging occurring at the time of the alarm? (Yes or (No))

2. Alarm caused by:

- | | |
|-------------------------------------|---|
| <input type="checkbox"/> | Loss of Power |
| <input type="checkbox"/> | Loss of Radio Communication |
| <input type="checkbox"/> | Out of Calibration |
| <input type="checkbox"/> | UV Alignment |
| <input type="checkbox"/> | Blockage in Air Tube |
| <input type="checkbox"/> | Air Quality |
| <input checked="" type="checkbox"/> | Other: weather, cold and windy, snowing |

3. Corrective Actions taken? (Yes or (No))

4. Dredging suspended? (Yes or (No))

5. Alarm logged in air monitoring action spreadsheet? ((Yes) or No)

Description of Action Taken:

A) No problems were found, snow blowing on top of the dike.