IHC CDF Dredge Project : Air Monitoring Alarm Response Log

Air Monitoring Station:	□ M1	□ M2	X M3	□ M4	□Dock
Air Monitoring Instrument	:				
☐ Cerex Unit (Naphthalene)	X	Thermo Unit (F	Particles)	☐ PID Monit	cor
Date & Time of alarm: 6/16/	2015 1:30AM	[Email sen	t to site Technician?	(Yes or No)
Technician responded to the alarm; Scott Peterson					
1. Was dredging occurring at the time of the alarm?					(Yes or No)
2. Alarm caused by:					
Loss of Pow	ver				
Loss of Radio Communication					
Out of Calibration					
UV Alignment					
Blockage in Air Tube					
Air Quality					
X Other: Insec	ets				
3. Corrective Actions taken?				(Yes or No	
4. Dredging suspended?					(Yes or No)
5. Alarm logged in air monitoring action spreadsheet?					(Yes) or No)

Description of Action Taken:

A) No problems found wrong with the air unit. Video camera records were reviewed, it was rainy and foggy at the time of the alarm. It is assume that a small spider was suck into the intake of the air unit.