

# IHC CDF Dredge Project : Air Monitoring Alarm Response Log

**Air Monitoring Station:**     M1             M2             M3             M4             Dock

**Air Monitoring Instrument:**

Cerex Unit (Naphthalene)             Thermo Unit (Particles)             PID Monitor

Date & Time of alarm: 9/10/2015 8:15AM & 11:30 PM            Email sent to site Technician?             Yes or  No

Technician responded to the alarm; Scott Peterson

1. Was dredging occurring at the time of the alarm?            ( Yes or  No )
2. Alarm caused by:

<input type="checkbox"/>	Loss of Power
<input type="checkbox"/>	Loss of Radio Communication
<input type="checkbox"/>	Out of Calibration
<input type="checkbox"/>	UV Alignment
<input type="checkbox"/>	Blockage in Air Tube
<input checked="" type="checkbox"/>	Air Quality
<input checked="" type="checkbox"/>	Other: Weather, Rain

3. Corrective Actions taken?            ( Yes or  No )
4. Dredging suspended?            ( Yes or  No )
5. Alarm logged in air monitoring action spreadsheet?            ( Yes or  No )

***Description of Action Taken:***

The reason for the air alarm is unknown. Air unit calibration was checked and is with-in tolerance.