\mathbf{X} M1 \square M2 \square M3 \square M4 □ Dock **Air Monitoring Station: Air Monitoring Instrument:** X Thermo Unit (Particles) ☐ PID Monitor ☐ Cerex Unit (Naphthalene) Date & Time of alarm: <u>6/10/2017 4:30 PM</u> Email sent to site Technician? (Yes)or No) Technician responded to the alarm; Scott Peterson (Yes or No 1. Was dredging occurring at the time of the alarm? 2. Alarm caused by: Loss of Power Loss of Radio Communication Out of Calibration **UV** Alignment Blockage in Air Tube Air Quality Other: weather, high winds 3. Corrective Actions taken? (Yes or (No Dredging suspended? (Yes or (No))

IHC CDF Dredge Project: Air Monitoring Alarm Response Log

(Yes) or No)

Description of Action Taken:

5. Alarm logged in air monitoring action spreadsheet?

A) No problems were found with air monitors. All air unit parameters were with-in tolorance during calibration check.