

IHC CDF Dredge Project : Air Monitoring Alarm Response Log

Air Monitoring Station: M1 M2 M3 M4 Dock

Air Monitoring Instrument:

Cerex Unit (Naphthalene) Thermo Unit (Particles) PID Monitor

Date & Time of alarm: 9/14/2017 to 9/17/2017

Email sent to site Technician? Yes or No

Technician responded to the alarm; Damian Allen

1. Was dredging occurring at the time of the alarm? (Yes or No)

2. Alarm caused by:

- | | |
|-------------------------------------|-----------------------------|
| <input type="checkbox"/> | Loss of Power |
| <input type="checkbox"/> | Loss of Radio Communication |
| <input type="checkbox"/> | Out of Calibration |
| <input type="checkbox"/> | UV Alignment |
| <input type="checkbox"/> | Blockage in Air Tube |
| <input type="checkbox"/> | Air Quality |
| <input checked="" type="checkbox"/> | Other: Spiders |

3. Corrective Actions taken? (Yes or No)

4. Dredging suspended? (Yes or No)

5. Alarm logged in air monitoring action spreadsheet? (Yes or No)

Description of Action Taken:

- A) M4 air unit parameters were within tolerance during calibration check. M1 & M4 were vacuumed and cleaned then calibrated.