	IC CDF Dre	edge Proje	ct : Air Mor	nitoring Alarn	n Response Log
Air Monitoring Station:	□ M1	□ м2	□ M3	X M4	Dock
Air Monitoring Instrument:□ Cerex Unit (Naphthalene)Image: Thermo Unit (Particles)				□ PID Monite	or
Date & Time of alarm:8/17 and 8/18Email sent to site Technician?					(Yesor No)
Technician responded to the a	alarm? Scott Pe	terson			
1. Was dredging occurring at the time of the alarm?					(Yes or No)
2. Alarm caused by:					
Loss of Pov	ver				
Loss of Rad	lio Communica	tion			
Out of Calibration					
UV Alignment					
X Blockage in Air Tube					
Air Quality					
X Other: unkn	iown, maybe ca	used by small in	nsects		
3. Corrective Actions taken?					(Vesor No)
4. Dredging suspended?					(Yes or No)
5. Alarm logged in air monitoring action spreadsheet?					(Yes) or No)

Description of Action Taken:

Air unit was re-calibrated. No problems were found during the inspections of the air units, vacuumed inlet tube going to filter paper roll.