IHC CDF Dredge Project: Air Monitoring Alarm Response Log

Air Monitoring	Station:	X M1	□ M2	□ M3	□ M4	□ Dock
Air Monitoring Instrument:						
☐ Cerex Unit (Naphthalene) ☐ Thermo Unit (Parti				articles)	□ PID Mon	nitor
Date & Time of alarm: <u>8/31/2020 7:45 PM</u>				7: Email se	nt to site Technic	ian? (Yes or No)
Technician responded to the alarm? Scott Peterson						
1. Was dredging occurring at the time of the alarm?						(Yes or No)
2. Alarm caused by:						
	Loss of Powe	er				
	Loss of Radio Communication					
	Out of Calibr	ration				
	UV Alignmen	nt				
Blockage in Air Tube						
	Air Quality					
X	Other: Unkno	own				
3. Corrective Actions taken?						(Yes or No)
4. Dredging suspended?						(Yes or No)
5. Alarm le	5. Alarm logged in air monitoring action spreadsheet?					(Yes) or No)

Description of Action Taken:

No problems air monitor inspection. Weather may have contributed to the high particles' readings. Air unit was recalibrated on 9/01/2020.