## IHC CDF Dredge Project: Air Monitoring Alarm Response Log

Air Monitoring Station:	□ M1	☑ M2	□ M3	□ M4	□ Dock
<b>Air Monitoring Instrumen</b>	nt:				
☐ Cerex Unit (Naphthalene)			Particles)	☐ PID Mor	nitor
Date & Time of alarm: 9/1/2	7: Email se	nt to site Technic	ian? (Yes or No)		
Technician responded to the alarm? Scott Peterson					
1. Was dredging occurring at the time of the alarm?					(Yes or No)
2. Alarm caused by:					
Loss of Po	ower				
Loss of R	Loss of Radio Communication				
Out of Ca	Out of Calibration				
UV Align	UV Alignment				
Blockage in Air Tube					
Air Quality					
X Other: Un	known				
3. Corrective Actions taken?					(Yes or No)
4. Dredging suspended?					(Yes or No)
5. Alarm logged in air monitoring action spreadsheet?				(Yes) or No)	

## Description of Action Taken:

No problems during air monitor inspection. Weather may have contributed to the high particles' readings. Air unit was recalibrated on 9/02/2020.