IHC CDF Dredge Project : Air Monitoring Alarm Response Log

Air Mo	onitoring	Station:	X M1	☑ M2	□ M3	□ M4	□ Dock
Air Monitoring Instrument:							
☐ Cerex Unit (Naphthalene) ☐ Thermo Unit (Part					articles)	☐ PID Monito	or
Date & Time of alarm: <u>2/6/21 8:45 PM</u>					Email sent	to site Technician?	(Yes or No)
Technician responded to the alarm? Scott Peterson							
1.	1. Was dredging occurring at the time of the alarm?						(Yes or No)
2.	2. Alarm caused by:						
		Loss of Powe	er				
	Loss of Radio Communication						
		Out of Calibr	ration				
		UV Alignme	nt				
		Blockage in A	Air Tube				
	X	Air Quality					
	X	Other: Weath	ner				
3.	. Corrective Actions taken?						(Yes or No)
4.	Dredging suspended?						(Yes or No)
5.	Alarm logged in air monitoring action spreadsheet?						(Yes)or No)

Description of Action Taken:

A) High winds and rains lead to the air alarms. M1 & M2 air unit was checked and recalibrated.