## IHC CDF Dredge Project: Air Monitoring Alarm Response Log

Air Monitoring Station:	□ M1	<b>■</b> M2	□ M3	□ M4	□ Dock
Air Monitoring Instrument:					
☐ Cerex Unit (Naphthalene)	$\mathbf{x}$	Thermo Unit (Pa	articles)	☐ PID Monito	or
Date & Time of alarm: 7/26 to	7/28		Email sent	to site Technician?	(Yes or No)
Technician responded to the alarm? Scott Peterson					
1. Was dredging occurring at the time of the alarm?					(Yes or No)
2. Alarm caused by:					
X Loss of Pow	er				
Loss of Rad	io Communic	ation			
X Out of Calib	X Out of Calibration				
UV Alignment					
X Blockage in Air Tube					
Air Quality					
X Other: Unkr	nown				
3. Corrective Actions taken?				(Yesor No)	
4. Dredging suspended?					(Yes or No)
5. Alarm logged in air monitoring action spreadsheet?					(Yes)or No)

## Description of Action Taken:

A) Air unit was checked several times and recalibrated. Found water from rainstorm leaked into air unit and dripped on the filter paper. Air unit was cleanup/ dried out and recalibrated.