

IHC CDF Dredge Project : Air Monitoring Alarm Response Log

Air Monitoring Station: M1 M2 M3 M4 Dock

Air Monitoring Instrument:

Cerex Unit (Naphthalene) Thermo Unit (Particles) PID Monitor

Date & Time of alarm: 12/16 6:00 AM

Email sent to site Technician? (Yes or No)

Technician responded to the alarm? Scott Peterson

1. Was dredging occurring at the time of the alarm? (Yes or No)

2. Alarm caused by:

- | | |
|-------------------------------------|-----------------------------|
| <input type="checkbox"/> | Loss of Power |
| <input type="checkbox"/> | Loss of Radio Communication |
| <input type="checkbox"/> | Out of Calibration |
| <input type="checkbox"/> | UV Alignment |
| <input type="checkbox"/> | Blockage in Air Tube |
| <input type="checkbox"/> | Air Quality |
| <input checked="" type="checkbox"/> | Other: Weather |

3. Corrective Actions taken? (Yes or No)

4. Dredging suspended? (Yes or No)

5. Alarm logged in air monitoring action spreadsheet? (Yes or No)

Description of Action Taken:

A) Cold weather may of caused the air units to alarm. Air units were checked and recalibrated on December 17th.