

IHC CDF Dredge Project : Air Monitoring Alarm Response Log

Air Monitoring Station: M1 M2 M3 M4 Dock

Air Monitoring Instrument:

Cerex Unit (Naphthalene) Thermo Unit (Particles) PID Monitor

Date & Time of alarm: 04/19/22 7:15 AM- 11:00 PM

Email sent to site Technician? (Yes or No)

Technician responded to the alarm? Scott Peterson

1. Was dredging or dike construction occurring at the time of the alarm? (Yes or No)

2. Alarm caused by:

- | | |
|---------------------------------------|-----------------------------|
| <input type="checkbox"/> | Loss of Power |
| <input type="checkbox"/> | Loss of Radio Communication |
| <input type="checkbox"/> | Out of Calibration |
| <input type="checkbox"/> | UV Alignment |
| <input type="checkbox"/> | Blockage in Air Tube |
| <input type="checkbox"/> ? | Air Quality, truck traffic |
| <input checked="" type="checkbox"/> X | Other: Unknown |

3. Corrective Actions taken? (Yes or No)

4. Dredging suspended? No dredging operation currently on site. (Yes or No)

5. Alarm logged in air monitoring action spreadsheet? (Yes) or No)

Description of Action Taken:

A) . It is unknown what is caused the high particle reads. The air unit was checked and calibrated. No deficiencies were found.