

# IHC CDF Dredge Project : Air Monitoring Alarm Response Log

**Air Monitoring Station:**     M1             M2             M3             M4             Dock

**Air Monitoring Instrument:**

Cerex Unit (Naphthalene)             Thermo Unit (Particles)             PID Monitor

Date & Time of alarm: 05/3/22

Email sent to site Technician? (Yes or **No**)

Technician responded to the alarm? Scott Peterson

1. Was dredging or dike construction occurring at the time of the alarm? (Yes or **No**)

2. Alarm caused by:

- |                                     |                             |
|-------------------------------------|-----------------------------|
| <input type="checkbox"/>            | Loss of Power               |
| <input type="checkbox"/>            | Loss of Radio Communication |
| <input type="checkbox"/>            | Out of Calibration          |
| <input type="checkbox"/>            | UV Alignment                |
| <input type="checkbox"/>            | Blockage in Air Tube        |
| <input type="checkbox"/>            | Air Quality, truck traffic  |
| <input checked="" type="checkbox"/> | Other: Unknown              |

3. Corrective Actions taken? ( **Yes** or No )

4. Dredging suspended? No dredging operation currently on site. ( Yes or **No** )

5. Alarm logged in air monitoring action spreadsheet? (**Yes** or No )

***Description of Action Taken:***

A) . It is unknown what is caused the high particle reads. The air unit was checked and calibrated. No deficiencies were found.