## IHC CDF Dredge Project : Air Monitoring Alarm Response Log

Air Monitoring Station:	⊠ M1	□ M2	□ M3	□ M4	□ Dock
Air Monitoring Instrumen	nt:				
☐ Cerex Unit (Naphthalene)			□ PID M	Ionitor	
Date & Time of alarm: <u>08/22/2022</u>					
Email sent to site Technician? (Yes)or No)			)		
Technician responded to the alarm? Scott Peterson					
1. Was dredging or dike construction occurring at the time of the alarm?					(Yes or No)
2. Alarm caused by:					
Loss of Power					
Loss of Radio Communication					
X Out of Ca	X Out of Calibration				
UV Align	UV Alignment				
Blockage in Air Tube					
Air Quali	Air Quality				
Other: filt	ter tape, power o	utage.			
					_
3. Corrective Actions taken?					(Yesor No)
4. Dredging suspended? No dredging operation currently on site.					(Yes or No)
5. Alarm logged in air monitoring action spreadsheet?					(Yes) or No)

## Description of Action Taken:

A) Lost site power during the night. Recalibrated the air unit M1.