## IHC CDF Dredge Project: Air Monitoring Alarm Response Log

Air Mon	nitoring	<b>Station:</b>	X M1	□ M2	□ M3	□ M4	□ Dock
Air Monitoring Instrument:							
☐ Cerex	x Unit (1	Naphthalene)	$\mathbf{x}$	Thermo Unit (Pa	articles)	□ PID Monitor	
Date & Time of alarm: 12/01/2022 9:00 AM-9:30 AM							
Email se	ent to site	e Technician?			(Yes)or No)		
Technician responded to the alarm? Scott Peterson							
1.	1. Was dredging or dike construction occurring at the time of the alarm?						(Yes or No)
2. Alarm caused by:							
		Loss of Powe	er				
		Loss of Radio	o Communic	ation			
		Out of Calibr	ration				
		UV Alignme	nt				
		Blockage in A	Air Tube				
	?	Air Quality, t	truck traffic,	Monday			
	X	Other: Unkno	own				
3. Corrective Actions taken?						(Yes or No)	
4. Dredging suspended? No dredging operation currently on site.							(Yes or No)
5. Alarm logged in air monitoring action spreadsheet?							(Yes) or No)

## Description of Action Taken:

- A) It is unknown what is causing the high particle reads. The air units were checked and calibrated. No deficiencies were found during inspection.
- B) Current construction with the dike project has increased semitruck traffic on the site delivering building materials. Although dust control measures are in place to spray the roads with water occasionally areas dry up allowing dust from traffic.
- C) Air units are checked and calibrated once a week.