## IHC CDF Dredge Project: Air Monitoring Alarm Response Log

Air Monito	ring Station:	X M1	□ M2	□ M3	□ M4	□ Dock
Air Monito	ring Instrument:					
☐ Cerex Unit (Naphthalene) ☐ Thermo Unit (Particles)					□ PID Mo	onitor
Date & Tim	e of alarm: <u>03/01/</u>	23 7:45AM				
Email sent to site Technician? (Yes)or No)					)	
Technician i	responded to the al	larm? Scott P	eterson			
1. Was dredging or dike construction occurring at the time of the alarm?						(Yes or No)
2. Alarm caused by:						
	Loss of Pow	er				
	Loss of Radio Communication					
	Out of Calib	ration				
	UV Alignme	ent				
Blockage in Air Tube						
	? Air Quality,	truck traffic,				
	X Other: Unkn	own				
3. Corrective Actions taken?						(Yes or No)
4. Dredging suspended? No dredging operation currently on site.						(Yes or No)
5. Alarm logged in air monitoring action spreadsheet?						(Yes) or No)

## Description of Action Taken:

- A) It is unknown what is causing the high particle reads. The air units were checked and calibrated. No deficiencies were found during inspection.
- B) Current construction with the dike project has increased semitruck traffic on the site delivering building materials. Although dust control measures are in place to spray the roads with water occasionally areas dry up allowing dust from traffic.
- C) Air units are checked and calibrated once a week.