IHC CDF Dredge Project: Air Monitoring Alarm Response Log

Air Monitorin	g Station:	X M1	X M2	□ M3	□ M 4	□ Dock
Air Monitoring Instrument:						
☐ Cerex Unit (Naphthalene)			Thermo Unit (Pa	articles)	☐ PID Monito	r
Date & Time of alarm: 06/18/2023 1:00PM-7:00PM5:30PM-6:30PM						
Email sent to site Technician?				(Yes)or No)		
Technician responded to the alarm? Scott Peterson						
1. Was dredging or dike construction occurring at the time of the alarm?						(Yes or No)
2. Alarm caused by:						
	Loss of Powe	er				
	Loss of Radi	o Communicat	tion			
	Out of Calib	ration				
	UV Alignme	nt				
Blockage in Air Tube						
?	? Air Quality					
X	Other: Unknown	own Weekend				
						_
3. Corrective Actions taken?					(Yesor No)	
4. Dredging suspended? No dredging operation currently on site.						(Yes or No)
5. Alarm logged in air monitoring action spreadsheet?						(Yes) or No)

Description of Action Taken:

A) It is unknown what caused the high particle reads. The air units were checked and calibrated. No deficiencies were found during inspection. Alarms occurred on the weekend or holiday while no activates or construction was occurring.