

IHC CDF Dredge Project : Air Monitoring Alarm Response Log

Air Monitoring Station: M1 M2 M3 M4 Dock

Air Monitoring Instrument:

Cerex Unit (Naphthalene) Thermo Unit (Particles) PID Monitor

Date & Time of alarm: 4/27/2013 8:00 PM Email sent to site Technician? (Yes or No)

Technician responded to the alarm; Scott Peterson

1. Was dredging occurring at the time of the alarm? (Yes or No)

2. Alarm caused by:

- | | |
|-------------------------------------|-----------------------------|
| <input type="checkbox"/> | Loss of Power |
| <input type="checkbox"/> | Loss of Radio Communication |
| <input type="checkbox"/> | Out of Calibration |
| <input type="checkbox"/> | UV Alignment |
| <input type="checkbox"/> | Blockage in Air Tube |
| <input checked="" type="checkbox"/> | Air Quality |
| <input checked="" type="checkbox"/> | Other: Weather |

3. Corrective Actions taken? (Yes or No)

4. Dredging suspended? (Yes or No)

5. Alarm logged in air monitoring action spreadsheet? (Yes or No)

Description of Action Taken:

Thermo units alarmed from high winds kicking up dust. Units were inspected at 9:30 PM 4/28/13 to confirm proper functions. Units were re-inspected at 7:30 AM 4/29/13 to confirm proper functions.