

IHC CDF Dredge Project : Air Monitoring Alarm Response Log

Air Monitoring Station: M1 M2 M3 M4 Dock

Air Monitoring Instrument:

Cerex Unit (Naphthalene) Thermo Unit (Particles) PID Monitor

Date & Time of alarm: 6/11/2014 12:30AM Email sent to site Technician? (Yes or No)

Technician responded to the alarm; Scott Peterson

1. Was dredging occurring at the time of the alarm? (Yes or No)

2. Alarm caused by:

- Loss of Power
- Loss of Radio Communication
- Out of Calibration
- UV Alignment
- Blockage in Air Tube
- Air Quality
- Other: Weather

3. Corrective Actions taken? (Yes or No)

4. Dredging suspended? (Yes or No)

5. Alarm logged in air monitoring action spreadsheet? (Yes or No)

Description of Action Taken:

- The surveillance cameras were checked to see if anything unusual around the site. Air units were checked and no problems were found. It was raining at the time of the alarm.