

IHC CDF Dredge Project : Air Monitoring Alarm Response Log

Air Monitoring Station: M1 M2 M3 M4 Dock

Air Monitoring Instrument:

Cerex Unit (Naphthalene) Thermo Unit (Particles) PID Monitor

Date & Time of alarm: 6/21/2014 4:30AM-7:00AM Email sent to site Technician? (Yes or No)

Technician responded to the alarm; Scott Peterson

1. Was dredging occurring at the time of the alarm? (Yes or No)

2. Alarm caused by:

- Loss of Power
- Loss of Radio Communication
- Out of Calibration
- UV Alignment
- Blockage in Air Tube
- Air Quality
- Other: Weather

3. Corrective Actions taken? (Yes or No)

4. Dredging suspended? (Yes or No)

5. Alarm logged in air monitoring action spreadsheet? (Yes or No)

Description of Action Taken:

- o Alarms were caused from thunderstorms and high winds. Unit was checked the next morning no problems were found.