

IHC CDF Dredge Project : Air Monitoring Alarm Response Log

Air Monitoring Station: M1 M2 M3 M4 Dock

Air Monitoring Instrument:

Cerex Unit (Naphthalene) Thermo Unit (Particles) PID Monitor

Date & Time of alarm: 6/28/2014 11:30 _____ Email sent to site Technician? ((Yes) or No)

Technician responded to the alarm; Scott Peterson

1. Was dredging occurring at the time of the alarm? ((Yes) or No)

2. Alarm caused by:

- Loss of Power
- Loss of Radio Communication
- Out of Calibration
- UV Alignment
- Blockage in Air Tube
- Air Quality
- Other: Spiders

3. Corrective Actions taken? (Yes or (No)

4. Dredging suspended? (Yes or (No)

5. Alarm logged in air monitoring action spreadsheet? ((Yes) or No)

Description of Action Taken:

- o Spiders inside of the nephelometer and paper filter caused M4 Thermo Unit alarm. The spiders were vacuumed out of the inlet tube and nephelometer. Unit was recalibrated at 8:30 AM 6/28/14 and inspected at 3:00 PM 6/28/14 to confirm proper function.

