

IHC CDF Dredge Project : Air Monitoring Alarm Response Log

Air Monitoring Station: M1 M2 M3 M4 Dock

Air Monitoring Instrument:

Cerex Unit (Naphthalene) Thermo Unit (Particles) PID Monitor

Date & Time of alarm: 10/05/20 All day Email sent to site Technician? (Yes or No)

Technician responded to the alarm? Scott Peterson

- 1. Was dredging occurring at the time of the alarm? (Yes or No)
- 2. Alarm caused by:

<input type="checkbox"/>	Loss of Power
<input type="checkbox"/>	Loss of Radio Communication
<input checked="" type="checkbox"/>	Out of Calibration
<input type="checkbox"/>	UV Alignment
<input type="checkbox"/>	Blockage in Air Tube
<input type="checkbox"/>	Air Quality
<input type="checkbox"/>	Other:

- 3. Corrective Actions taken? (Yes or No)
- 4. Dredging suspended? (Yes or No)
- 5. Alarm logged in air monitoring action spreadsheet? (Yes or No)

Description of Action Taken:

A) Alarms cause because air unit went out of calibration. M4 air unit was checked the next morning and recalibrated.