

# IHC CDF Dredge Project : Air Monitoring Alarm Response Log

**Air Monitoring Station:**     M1             M2             M3             M4             Dock

**Air Monitoring Instrument:**

Cerex Unit (Naphthalene)             Thermo Unit (Particles)             PID Monitor

Date & Time of alarm: 11/01/2020 3:45 PM

7: Email sent to site Technician?    (Yes or **No**)

Technician responded to the alarm? Scott Peterson

1. Was dredging occurring at the time of the alarm?            ( Yes or **No**)

2. Alarm caused by:

- |                                     |                             |
|-------------------------------------|-----------------------------|
| <input type="checkbox"/>            | Loss of Power               |
| <input type="checkbox"/>            | Loss of Radio Communication |
| <input type="checkbox"/>            | Out of Calibration          |
| <input type="checkbox"/>            | UV Alignment                |
| <input type="checkbox"/>            | Blockage in Air Tube        |
| <input type="checkbox"/>            | Air Quality                 |
| <input checked="" type="checkbox"/> | Other: Unknown              |

3. Corrective Actions taken?            ( Yes or **No**)

4. Dredging suspended?            ( Yes or **No**)

5. Alarm logged in air monitoring action spreadsheet?            (**Yes** or No )

***Description of Action Taken:***

No problems noted during air monitor inspection. Weather may have contributed to the high particles' readings.  
Air unit was recalibrated on 11/2/2020.